

Calibration Results

Probe file=5X50 Date=9/26/2016 Time=7:51:47 AM

T1A0B0 ERROR X 0.0000 Y 0.0000 Z 0.0000 D= 25.0089 SD= 0.0557 PR 0.0000

Max X= 0.0000 Y= 0.0000 Z= 0.0000 PR= 0.0000

Min X= 0.0000 Y= 0.0000 Z= 0.0000

NUMBER MEASURED=1


cal_sphere CENT X 189.1232 Y 308.1153 Z-452.0730 D24.9824

Standard deviations for probe calibration exceed limit.

T1A0B0

Measured probe diameter error exceeds the limits.

T1A0B0



Copy

Print

OK

Cancel